



October 17, 2023

Andrew Schuck
Dundee Central School District
55 Water Street

Dundee, NY 14837

RE: Project: LEAD TESTING 10/4
Pace Project No.: 70273383

Dear Andrew Schuck:

Enclosed are the analytical results for sample(s) received by the laboratory on October 09, 2023. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Melville

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

A handwritten signature in cursive script that reads "Daniel H. Bonitto".

Daniel H. Bonitto
daniel.bonitto@pacelabs.com
516-370-6000
Project Manager

Enclosures



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CERTIFICATIONS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Pace Analytical Services Long Island

575 Broad Hollow Rd, Melville, NY 11747

Connecticut Certification #: PH-0435

Delaware Certification # NY 10478

Maryland Certification #: 208

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

New Jersey Certification #: NY158

New York Certification #: 10478 Primary Accrediting Body

Pennsylvania Certification #: 68-00350

Rhode Island Certification #: LAO00340

Virginia Certification # 460302

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 1		Lab ID: 70273383001	Collected: 10/04/23 03:25	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	7.0	ug/L	1.0	1		10/13/23 10:54	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 2		Lab ID: 70273383002	Collected: 10/04/23 03:25	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	12.5	ug/L	1.0	1		10/13/23 10:59	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 3		Lab ID: 70273383003	Collected: 10/04/23 03:29	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 11:03	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 4		Lab ID: 70273383004	Collected: 10/04/23 03:29	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 11:05	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 5		Lab ID: 70273383005	Collected: 10/04/23 03:32	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.5	ug/L	1.0	1		10/13/23 11:06	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 6		Lab ID: 70273383006	Collected: 10/04/23 03:32	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.3	ug/L	1.0	1		10/13/23 11:08	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 7		Lab ID: 70273383007	Collected: 10/04/23 03:33	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.1	ug/L	1.0	1		10/13/23 11:10	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 8		Lab ID: 70273383008	Collected: 10/04/23 03:33	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	4.7	ug/L	1.0	1		10/13/23 11:11	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 9		Lab ID: 70273383009	Collected: 10/04/23 03:35	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.6	ug/L	1.0	1		10/13/23 11:13	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 10		Lab ID: 70273383010	Collected: 10/04/23 03:36	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.7	ug/L	1.0	1		10/13/23 11:17	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 11		Lab ID: 70273383011	Collected: 10/04/23 03:36	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.7	ug/L	1.0	1		10/13/23 11:25	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 12		Lab ID: 70273383012	Collected: 10/04/23 03:37	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 11:29	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 13		Lab ID: 70273383013	Collected: 10/04/23 03:37	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 11:31	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 14		Lab ID: 70273383014	Collected: 10/04/23 03:39	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.4	ug/L	1.0	1		10/13/23 11:32	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 15		Lab ID: 70273383015		Collected: 10/04/23 03:39	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.6	ug/L	1.0	1		10/13/23 11:34	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 16		Lab ID: 70273383016	Collected: 10/04/23 03:42	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	18.4	ug/L	1.0	1		10/13/23 11:36	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 17		Lab ID: 70273383017	Collected: 10/04/23 03:43	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	6.2	ug/L	1.0	1		10/13/23 11:40	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 18		Lab ID: 70273383018		Collected: 10/04/23 03:45	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.1	ug/L	1.0	1		10/13/23 11:42	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 19		Lab ID: 70273383019	Collected: 10/04/23 03:46	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		10/13/23 11:43	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 20		Lab ID: 70273383020	Collected: 10/04/23 03:51	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	13.5	ug/L	1.0	1		10/13/23 11:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 21		Lab ID: 70273383021	Collected: 10/04/23 03:52	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.4	ug/L	1.0	1		10/13/23 11:46	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 22		Lab ID: 70273383022	Collected: 10/04/23 03:52	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.5	ug/L	1.0	1		10/13/23 11:48	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 23		Lab ID: 70273383023	Collected: 10/04/23 03:55	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 11:50	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 24		Lab ID: 70273383024		Collected: 10/04/23 03:56	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.1	ug/L	1.0	1		10/13/23 11:51	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 25		Lab ID: 70273383025	Collected: 10/04/23 03:56	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 11:53	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 26		Lab ID: 70273383026		Collected: 10/04/23 03:58	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	7.6	ug/L	1.0	1		10/13/23 11:54	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 27		Lab ID: 70273383027	Collected: 10/04/23 03:59	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.9	ug/L	1.0	1		10/13/23 11:59	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 28		Lab ID: 70273383028	Collected: 10/04/23 03:59	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.3	ug/L	1.0	1		10/13/23 12:00	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 30		Lab ID: 70273383029	Collected: 10/04/23 04:03	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.8	ug/L	1.0	1		10/13/23 12:02	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 31		Lab ID: 70273383030	Collected: 10/04/23 04:04	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.3	ug/L	1.0	1		10/13/23 12:07	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 32		Lab ID: 70273383031		Collected: 10/04/23 04:04	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.4	ug/L	1.0	1		10/13/23 12:11	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 33		Lab ID: 70273383032	Collected: 10/04/23 04:05	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	13.2	ug/L	1.0	1		10/13/23 12:19	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 34		Lab ID: 70273383033	Collected: 10/04/23 04:07	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 12:20	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 35		Lab ID: 70273383034	Collected: 10/04/23 04:07	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 12:22	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 36		Lab ID: 70273383035	Collected: 10/04/23 04:10	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 12:23	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 37		Lab ID: 70273383036	Collected: 10/04/23 04:10	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 12:25	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 38		Lab ID: 70273383037		Collected: 10/04/23 04:12		Received: 10/09/23 09:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/13/23 12:26	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 39		Lab ID: 70273383038		Collected: 10/04/23 04:12		Received: 10/09/23 09:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.9	ug/L	1.0	1		10/13/23 12:28	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 40		Lab ID: 70273383039		Collected: 10/04/23 04:14	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		10/13/23 12:30	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 41		Lab ID: 70273383040		Collected: 10/04/23 04:14	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 12:31	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 42		Lab ID: 70273383041		Collected: 10/04/23 04:15	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 12:36	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 43		Lab ID: 70273383042		Collected: 10/04/23 04:16	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.2	ug/L	1.0	1		10/13/23 12:37	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 44		Lab ID: 70273383043	Collected: 10/04/23 04:17	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.5	ug/L	1.0	1		10/13/23 12:39	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 45		Lab ID: 70273383044	Collected: 10/04/23 04:17	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 12:40	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 46		Lab ID: 70273383045	Collected: 10/04/23 04:17	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 12:42	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 47		Lab ID: 70273383046	Collected: 10/04/23 04:18	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 12:43	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 48		Lab ID: 70273383047		Collected: 10/04/23 04:22		Received: 10/09/23 09:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	75.4	ug/L	1.0	1		10/13/23 12:45	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 49		Lab ID: 70273383048		Collected: 10/04/23 04:22		Received: 10/09/23 09:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	25.9	ug/L	1.0	1		10/13/23 12:46	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 50		Lab ID: 70273383049	Collected: 10/04/23 04:22	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.0	ug/L	1.0	1		10/13/23 12:48	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 51		Lab ID: 70273383050		Collected: 10/04/23 04:22		Received: 10/09/23 09:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	27.9	ug/L	1.0	1		10/13/23 12:56	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 52		Lab ID: 70273383051		Collected: 10/04/23 04:24	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.8	ug/L	1.0	1		10/13/23 13:00	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 53		Lab ID: 70273383052	Collected: 10/04/23 04:24	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	261	ug/L	1.0	1		10/13/23 13:05	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 54		Lab ID: 70273383053	Collected: 10/04/23 04:26	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.3	ug/L	1.0	1		10/13/23 13:06	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 55		Lab ID: 70273383054		Collected: 10/04/23 04:26	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.2	ug/L	1.0	1		10/13/23 13:08	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 56		Lab ID: 70273383055	Collected: 10/04/23 04:27	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.8	ug/L	1.0	1		10/13/23 13:12	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 57		Lab ID: 70273383056	Collected: 10/04/23 04:27	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	46.4	ug/L	1.0	1		10/13/23 13:14	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 58		Lab ID: 70273383057		Collected: 10/04/23 04:28	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	4.3	ug/L	1.0	1		10/13/23 13:15	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 59		Lab ID: 70273383058		Collected: 10/04/23 04:28		Received: 10/09/23 09:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	18.6	ug/L	1.0	1		10/13/23 13:17	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 60		Lab ID: 70273383059	Collected: 10/04/23 04:29	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 13:19	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 61		Lab ID: 70273383060	Collected: 10/04/23 04:29	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 13:20	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 62		Lab ID: 70273383061	Collected: 10/04/23 04:30	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.3	ug/L	1.0	1		10/13/23 13:22	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 63		Lab ID: 70273383062	Collected: 10/04/23 04:30	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	53.8	ug/L	1.0	1		10/13/23 13:23	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 64		Lab ID: 70273383063		Collected: 10/04/23 04:31	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.7	ug/L	1.0	1		10/13/23 13:25	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 65		Lab ID: 70273383064	Collected: 10/04/23 04:31	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	8.7	ug/L	1.0	1		10/13/23 13:26	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 66		Lab ID: 70273383065	Collected: 10/04/23 04:33	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	4.2	ug/L	1.0	1		10/13/23 13:31	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 68		Lab ID: 70273383066		Collected: 10/04/23 04:34	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.3	ug/L	1.0	1		10/13/23 13:32	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 70		Lab ID: 70273383067	Collected: 10/04/23 04:36	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.5	ug/L	1.0	1		10/13/23 13:34	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 71		Lab ID: 70273383068	Collected: 10/04/23 04:36	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	4.3	ug/L	1.0	1		10/13/23 13:35	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 72		Lab ID: 70273383069	Collected: 10/04/23 04:37	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.8	ug/L	1.0	1		10/13/23 13:37	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 73		Lab ID: 70273383070	Collected: 10/04/23 04:38	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.5	ug/L	1.0	1		10/13/23 13:42	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 74		Lab ID: 70273383071		Collected: 10/04/23 04:38	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.1	ug/L	1.0	1		10/13/23 13:49	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 75		Lab ID: 70273383072		Collected: 10/04/23 04:39	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.5	ug/L	1.0	1		10/13/23 13:54	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 76		Lab ID: 70273383073	Collected: 10/04/23 04:40	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.6	ug/L	1.0	1		10/13/23 13:55	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 77		Lab ID: 70273383074	Collected: 10/04/23 04:40	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	4.0	ug/L	1.0	1		10/13/23 13:57	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 78		Lab ID: 70273383075		Collected: 10/04/23 04:41	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.4	ug/L	1.0	1		10/13/23 13:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 79		Lab ID: 70273383076	Collected: 10/04/23 04:41	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	8.5	ug/L	1.0	1		10/13/23 14:00	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 80		Lab ID: 70273383077		Collected: 10/04/23 04:42	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	20.4	ug/L	1.0	1		10/13/23 14:01	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 81		Lab ID: 70273383078		Collected: 10/04/23 04:42	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	26.4	ug/L	1.0	1		10/13/23 14:03	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 82		Lab ID: 70273383079		Collected: 10/04/23 04:44	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 14:07	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 83		Lab ID: 70273383080	Collected: 10/04/23 04:44	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 14:09	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 84		Lab ID: 70273383081	Collected: 10/04/23 04:47	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 14:11	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 85		Lab ID: 70273383082	Collected: 10/04/23 04:47	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 14:12	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 86		Lab ID: 70273383083	Collected: 10/04/23 04:47	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 14:14	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 87		Lab ID: 70273383084	Collected: 10/04/23 04:50	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		10/13/23 14:15	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 88		Lab ID: 70273383085	Collected: 10/04/23 04:52	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 14:17	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 89		Lab ID: 70273383086		Collected: 10/04/23 04:52	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.1	ug/L	1.0	1		10/13/23 14:18	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 90		Lab ID: 70273383087	Collected: 10/04/23 04:55	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	33.1	ug/L	1.0	1		10/13/23 14:20	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 91		Lab ID: 70273383088	Collected: 10/04/23 04:55	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	40.3	ug/L	1.0	1		10/13/23 14:21	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 92		Lab ID: 70273383089		Collected: 10/04/23 04:58	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		10/13/23 14:26	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 93		Lab ID: 70273383090	Collected: 10/04/23 04:58	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.6	ug/L	1.0	1		10/13/23 14:31	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 94		Lab ID: 70273383091	Collected: 10/04/23 04:58	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.7	ug/L	1.0	1		10/13/23 14:35	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 95		Lab ID: 70273383092		Collected: 10/04/23 05:50	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 14:40	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 96		Lab ID: 70273383093	Collected: 10/04/23 05:50	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 14:44	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 97		Lab ID: 70273383094	Collected: 10/04/23 05:51	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	5.7	ug/L	1.0	1		10/13/23 14:46	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 98		Lab ID: 70273383095		Collected: 10/04/23 05:52	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.7	ug/L	1.0	1		10/13/23 14:48	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 99		Lab ID: 70273383096	Collected: 10/04/23 05:53	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	8.6	ug/L	1.0	1		10/13/23 14:49	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 100		Lab ID: 70273383097	Collected: 10/04/23 06:05	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	4.8	ug/L	1.0	1		10/13/23 14:51	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 101		Lab ID: 70273383098		Collected: 10/04/23 06:08	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Preparation Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1	10/13/23 07:18	10/13/23 17:56	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 102		Lab ID: 70273383099	Collected: 10/04/23 05:07	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		10/13/23 14:52	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 103		Lab ID: 70273383100		Collected: 10/04/23 05:07	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.7	ug/L	1.0	1		10/13/23 14:54	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 104		Lab ID: 70273383101	Collected: 10/04/23 05:09	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 14:55	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 105		Lab ID: 70273383102	Collected: 10/04/23 05:10	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.6	ug/L	1.0	1		10/13/23 14:57	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 107		Lab ID: 70273383103	Collected: 10/04/23 05:11	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 14:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 108		Lab ID: 70273383104	Collected: 10/04/23 05:11	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.1	ug/L	1.0	1		10/13/23 15:03	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 109		Lab ID: 70273383105		Collected: 10/04/23 05:12	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	6.2	ug/L	1.0	1		10/13/23 15:05	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 110		Lab ID: 70273383106	Collected: 10/04/23 05:12	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	10.7	ug/L	1.0	1		10/13/23 15:06	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 111		Lab ID: 70273383107	Collected: 10/04/23 05:12	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.6	ug/L	1.0	1		10/13/23 15:08	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 112		Lab ID: 70273383108	Collected: 10/04/23 05:12	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.8	ug/L	1.0	1		10/13/23 15:09	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 113		Lab ID: 70273383109	Collected: 10/04/23 05:12	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	39.9	ug/L	1.0	1		10/13/23 15:11	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 114		Lab ID: 70273383110	Collected: 10/04/23 05:15	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	5.5	ug/L	1.0	1		10/13/23 15:12	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 115		Lab ID: 70273383111	Collected: 10/04/23 05:15	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	6.4	ug/L	1.0	1		10/13/23 15:17	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 116		Lab ID: 70273383112	Collected: 10/04/23 05:15	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	8.2	ug/L	1.0	1		10/13/23 15:25	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 117		Lab ID: 70273383113	Collected: 10/04/23 05:15	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.6	ug/L	1.0	1		10/13/23 15:29	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 119		Lab ID: 70273383114	Collected: 10/04/23 05:19	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 15:31	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 120		Lab ID: 70273383115		Collected: 10/04/23 05:19	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.4	ug/L	1.0	1		10/13/23 15:32	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 121		Lab ID: 70273383116	Collected: 10/04/23 05:20	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		10/13/23 15:34	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 122		Lab ID: 70273383117		Collected: 10/04/23 05:21		Received: 10/09/23 09:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/13/23 15:35	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 123		Lab ID: 70273383118		Collected: 10/04/23 05:21	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.1	ug/L	1.0	1		10/13/23 15:40	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 124		Lab ID: 70273383119		Collected: 10/04/23 05:23		Received: 10/09/23 09:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	3.6	ug/L	1.0	1		10/13/23 15:41	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 125		Lab ID: 70273383120	Collected: 10/04/23 05:23	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.9	ug/L	1.0	1		10/13/23 15:43	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 126		Lab ID: 70273383121		Collected: 10/04/23 05:25	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.7	ug/L	1.0	1		10/13/23 15:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 127		Lab ID: 70273383122	Collected: 10/04/23 05:25	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	5.4	ug/L	1.0	1		10/13/23 15:46	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 128		Lab ID: 70273383123	Collected: 10/04/23 05:27	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.8	ug/L	1.0	1		10/13/23 15:48	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 129		Lab ID: 70273383124	Collected: 10/04/23 05:27	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	12.5	ug/L	1.0	1		10/13/23 15:49	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 130		Lab ID: 70273383125	Collected: 10/04/23 05:27	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	162	ug/L	1.0	1		10/13/23 15:51	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 131		Lab ID: 70273383126		Collected: 10/04/23 05:28	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	4.5	ug/L	1.0	1		10/13/23 15:52	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 132		Lab ID: 70273383127		Collected: 10/04/23 05:28	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Preparation Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	196	ug/L	1.0	1	10/13/23 07:18	10/13/23 17:57	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 133		Lab ID: 70273383128	Collected: 10/04/23 05:30	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.4	ug/L	1.0	1		10/13/23 15:54	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 134		Lab ID: 70273383129	Collected: 10/04/23 05:30	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.3	ug/L	1.0	1		10/13/23 15:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 135		Lab ID: 70273383130	Collected: 10/04/23 05:31	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.2	ug/L	1.0	1		10/13/23 16:00	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 136		Lab ID: 70273383131		Collected: 10/04/23 05:31	Received: 10/09/23 09:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.2	ug/L	1.0	1		10/13/23 16:01	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 137		Lab ID: 70273383132	Collected: 10/04/23 05:32	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.6	ug/L	1.0	1		10/13/23 16:06	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 138		Lab ID: 70273383133	Collected: 10/04/23 05:33	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	4.8	ug/L	1.0	1		10/13/23 16:11	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 139		Lab ID: 70273383134	Collected: 10/04/23 05:33	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	9.6	ug/L	1.0	1		10/13/23 16:18	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 140		Lab ID: 70273383135	Collected: 10/04/23 05:34	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.0	ug/L	1.0	1		10/13/23 16:20	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 141		Lab ID: 70273383136	Collected: 10/04/23 05:34	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	9.6	ug/L	1.0	1		10/13/23 16:21	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 142		Lab ID: 70273383137	Collected: 10/04/23 05:35	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	8.3	ug/L	1.0	1		10/13/23 16:23	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 143		Lab ID: 70273383138	Collected: 10/04/23 05:35	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	11.1	ug/L	1.0	1		10/13/23 16:25	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 144		Lab ID: 70273383139	Collected: 10/04/23 05:36	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	13.3	ug/L	1.0	1		10/13/23 16:26	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 145		Lab ID: 70273383140	Collected: 10/04/23 05:37	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	8.8	ug/L	1.0	1		10/13/23 16:28	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 146		Lab ID: 70273383141	Collected: 10/04/23 05:38	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	11.0	ug/L	1.0	1		10/13/23 16:29	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 147		Lab ID: 70273383142	Collected: 10/04/23 05:39	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.0	ug/L	1.0	1		10/13/23 16:31	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 148		Lab ID: 70273383143	Collected: 10/04/23 05:40	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	270	ug/L	1.0	1		10/13/23 16:35	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Sample: 69		Lab ID: 70273383144	Collected: 10/04/23 04:34	Received: 10/09/23 09:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.1	ug/L	1.0	1		10/13/23 16:37	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

QC Batch:	323806	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70273383001, 70273383002, 70273383003, 70273383004, 70273383005, 70273383006, 70273383007, 70273383008, 70273383009		

METHOD BLANK:	1653285	Matrix:	Water
Associated Lab Samples:	70273383001, 70273383002, 70273383003, 70273383004, 70273383005, 70273383006, 70273383007, 70273383008, 70273383009		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/13/23 10:26	

LABORATORY CONTROL SAMPLE:	1653286					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	51.6	103	85-115	

MATRIX SPIKE SAMPLE:	1653300						
Parameter	Units	70273635001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	23.0	50	68.3	91	70-130	

MATRIX SPIKE SAMPLE:	1653303						
Parameter	Units	70273383001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	7.0	50	54.8	96	70-130	

SAMPLE DUPLICATE:	1653299					
Parameter	Units	70273635001 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	23.0	116	134	D6	

SAMPLE DUPLICATE:	1653302					
Parameter	Units	70273383001 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	7.0	7.0	0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

QC Batch:	323807	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70273383010, 70273383011, 70273383012, 70273383013, 70273383014, 70273383015, 70273383016, 70273383017, 70273383018, 70273383019, 70273383020, 70273383021, 70273383022, 70273383023, 70273383024, 70273383025, 70273383026, 70273383027, 70273383028, 70273383029		

METHOD BLANK:	1653304	Matrix:	Water
Associated Lab Samples:	70273383010, 70273383011, 70273383012, 70273383013, 70273383014, 70273383015, 70273383016, 70273383017, 70273383018, 70273383019, 70273383020, 70273383021, 70273383022, 70273383023, 70273383024, 70273383025, 70273383026, 70273383027, 70273383028, 70273383029		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/13/23 11:14	

LABORATORY CONTROL SAMPLE:	1653305					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	55.4	111	85-115	

MATRIX SPIKE SAMPLE:	1653307						
Parameter	Units	70273383010 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	2.7	50	47.7	90	70-130	

MATRIX SPIKE SAMPLE:	1653309						
Parameter	Units	70273383011 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	3.7	50	51.4	95	70-130	

SAMPLE DUPLICATE:	1653306					
Parameter	Units	70273383010 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	2.7	2.6	1		

SAMPLE DUPLICATE:	1653308					
Parameter	Units	70273383011 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	3.7	3.7	1		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

QC Batch: 323808

Analysis Method: EPA 200.8

QC Batch Method: EPA 200.8

Analysis Description: 200.8 MET No Prep Drinking Water

Laboratory: Pace Analytical Services - Melville

Associated Lab Samples: 70273383030, 70273383031, 70273383032, 70273383033, 70273383034, 70273383035, 70273383036, 70273383037, 70273383038, 70273383039, 70273383040, 70273383041, 70273383042, 70273383043, 70273383044, 70273383045, 70273383046, 70273383047, 70273383048, 70273383049

METHOD BLANK: 1653310

Matrix: Water

Associated Lab Samples: 70273383030, 70273383031, 70273383032, 70273383033, 70273383034, 70273383035, 70273383036, 70273383037, 70273383038, 70273383039, 70273383040, 70273383041, 70273383042, 70273383043, 70273383044, 70273383045, 70273383046, 70273383047, 70273383048, 70273383049

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/13/23 12:03	

LABORATORY CONTROL SAMPLE: 1653311

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	51.3	103	85-115	

MATRIX SPIKE SAMPLE: 1653313

Parameter	Units	70273383030 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	2.3	50	50.0	95	70-130	

MATRIX SPIKE SAMPLE: 1653315

Parameter	Units	70273383031 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	2.4	50	59.4	114	70-130	

SAMPLE DUPLICATE: 1653312

Parameter	Units	70273383030 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	2.3	2.3	1	

SAMPLE DUPLICATE: 1653314

Parameter	Units	70273383031 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	2.4	2.4	0	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

QC Batch:	323809	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70273383050, 70273383051, 70273383052, 70273383053, 70273383054, 70273383055, 70273383056, 70273383057, 70273383058, 70273383059, 70273383060, 70273383061, 70273383062, 70273383063, 70273383064, 70273383065, 70273383066, 70273383067, 70273383068, 70273383069		

METHOD BLANK:	1653316	Matrix:	Water
Associated Lab Samples:	70273383050, 70273383051, 70273383052, 70273383053, 70273383054, 70273383055, 70273383056, 70273383057, 70273383058, 70273383059, 70273383060, 70273383061, 70273383062, 70273383063, 70273383064, 70273383065, 70273383066, 70273383067, 70273383068, 70273383069		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/13/23 12:50	

LABORATORY CONTROL SAMPLE:	1653317					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	51.3	103	85-115	

MATRIX SPIKE SAMPLE:	1653321						
Parameter	Units	70273383050 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	27.9	50	69.8	84	70-130	

MATRIX SPIKE SAMPLE:	1653323						
Parameter	Units	70273383051 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.8	50	47.9	92	70-130	

SAMPLE DUPLICATE:	1653320					
Parameter	Units	70273383050 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	27.9	28.1	1		

SAMPLE DUPLICATE:	1653322					
Parameter	Units	70273383051 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	1.8	1.8	1		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

QC Batch:	323810	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70273383070, 70273383071, 70273383072, 70273383073, 70273383074, 70273383075, 70273383076, 70273383077, 70273383078, 70273383079, 70273383080, 70273383081, 70273383082, 70273383083, 70273383084, 70273383085, 70273383086, 70273383087, 70273383088, 70273383089		

METHOD BLANK:	1653324	Matrix:	Water
Associated Lab Samples:	70273383070, 70273383071, 70273383072, 70273383073, 70273383074, 70273383075, 70273383076, 70273383077, 70273383078, 70273383079, 70273383080, 70273383081, 70273383082, 70273383083, 70273383084, 70273383085, 70273383086, 70273383087, 70273383088, 70273383089		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/13/23 13:39	

LABORATORY CONTROL SAMPLE:	1653325					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	51.5	103	85-115	

MATRIX SPIKE SAMPLE:	1653327						
Parameter	Units	70273383070 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	2.5	50	52.3	100	70-130	

MATRIX SPIKE SAMPLE:	1653329						
Parameter	Units	70273383071 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	3.1	50	51.2	96	70-130	

SAMPLE DUPLICATE:	1653326					
Parameter	Units	70273383070 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	2.5	2.5	1		

SAMPLE DUPLICATE:	1653328					
Parameter	Units	70273383071 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	3.1	3.1	1		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

QC Batch:	323811	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70273383090, 70273383091, 70273383092, 70273383093, 70273383094, 70273383095, 70273383096, 70273383097, 70273383099, 70273383100, 70273383101, 70273383102, 70273383103, 70273383104, 70273383105, 70273383106, 70273383107, 70273383108, 70273383109, 70273383110		

METHOD BLANK:	1653330	Matrix:	Water
Associated Lab Samples:	70273383090, 70273383091, 70273383092, 70273383093, 70273383094, 70273383095, 70273383096, 70273383097, 70273383099, 70273383100, 70273383101, 70273383102, 70273383103, 70273383104, 70273383105, 70273383106, 70273383107, 70273383108, 70273383109, 70273383110		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/13/23 14:28	

LABORATORY CONTROL SAMPLE:	1653331					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	49.7	99	85-115	

MATRIX SPIKE SAMPLE:	1653333						
Parameter	Units	70273383090 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.6	50	48.8	94	70-130	

MATRIX SPIKE SAMPLE:	1653335						
Parameter	Units	70273383091 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.7	50	49.1	95	70-130	

SAMPLE DUPLICATE:	1653332					
Parameter	Units	70273383090 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	1.6	1.6	4		

SAMPLE DUPLICATE:	1653334					
Parameter	Units	70273383091 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	1.7	1.7	2		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

QC Batch:	323812	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70273383111, 70273383112, 70273383113, 70273383114, 70273383115, 70273383116, 70273383117, 70273383118, 70273383119, 70273383120, 70273383121, 70273383122, 70273383123, 70273383124, 70273383125, 70273383126, 70273383128, 70273383129, 70273383130, 70273383131		

METHOD BLANK:	1653336	Matrix:	Water
Associated Lab Samples:	70273383111, 70273383112, 70273383113, 70273383114, 70273383115, 70273383116, 70273383117, 70273383118, 70273383119, 70273383120, 70273383121, 70273383122, 70273383123, 70273383124, 70273383125, 70273383126, 70273383128, 70273383129, 70273383130, 70273383131		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/13/23 15:14	

LABORATORY CONTROL SAMPLE:	1653337					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	48.2	96	85-115	

MATRIX SPIKE SAMPLE:	1653339						
Parameter	Units	70273383111 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	6.4	50	54.1	95	70-130	

MATRIX SPIKE SAMPLE:	1653341						
Parameter	Units	70273383112 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	8.2	50	55.2	94	70-130	

SAMPLE DUPLICATE:	1653338					
Parameter	Units	70273383111 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	6.4	6.4	2		

SAMPLE DUPLICATE:	1653340					
Parameter	Units	70273383112 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	8.2	8.1	1		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

QC Batch:	323813	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70273383132, 70273383133, 70273383134, 70273383135, 70273383136, 70273383137, 70273383138, 70273383139, 70273383140, 70273383141, 70273383142, 70273383143, 70273383144		

METHOD BLANK:	1653342	Matrix:	Water
Associated Lab Samples:	70273383132, 70273383133, 70273383134, 70273383135, 70273383136, 70273383137, 70273383138, 70273383139, 70273383140, 70273383141, 70273383142, 70273383143, 70273383144		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/13/23 16:03	

LABORATORY CONTROL SAMPLE: 1653343						
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	48.0	96	85-115	

MATRIX SPIKE SAMPLE: 1653345							
Parameter	Units	70273383132 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	3.6	50	50.2	93	70-130	

MATRIX SPIKE SAMPLE: 1653347							
Parameter	Units	70273383133 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	4.8	50	51.2	93	70-130	

SAMPLE DUPLICATE: 1653344					
Parameter	Units	70273383132 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	3.6	3.5	1	

SAMPLE DUPLICATE: 1653346					
Parameter	Units	70273383133 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	4.8	4.8	1	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

QC Batch:	323798	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET Drinking Water
		Laboratory:	Pace Analytical Services - Melville

Associated Lab Samples: 70273383098, 70273383127

METHOD BLANK: 1653252 Matrix: Water

Associated Lab Samples: 70273383098, 70273383127

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/13/23 17:40	

LABORATORY CONTROL SAMPLE: 1653253

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	52.3	105	85-115	

MATRIX SPIKE SAMPLE: 1653255

Parameter	Units	70273326076 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	5.0	50	54.6	99	70-130	

SAMPLE DUPLICATE: 1653254

Parameter	Units	70273326076 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	5.0	5.0	1	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALIFIERS

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

ANALYTE QUALIFIERS

D6 The precision between the sample and sample duplicate exceeded laboratory control limits.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70273383098	101	EPA 200.8	323798	EPA 200.8	323829
70273383127	132	EPA 200.8	323798	EPA 200.8	323829
70273383001	1	EPA 200.8	323806		
70273383002	2	EPA 200.8	323806		
70273383003	3	EPA 200.8	323806		
70273383004	4	EPA 200.8	323806		
70273383005	5	EPA 200.8	323806		
70273383006	6	EPA 200.8	323806		
70273383007	7	EPA 200.8	323806		
70273383008	8	EPA 200.8	323806		
70273383009	9	EPA 200.8	323806		
70273383010	10	EPA 200.8	323807		
70273383011	11	EPA 200.8	323807		
70273383012	12	EPA 200.8	323807		
70273383013	13	EPA 200.8	323807		
70273383014	14	EPA 200.8	323807		
70273383015	15	EPA 200.8	323807		
70273383016	16	EPA 200.8	323807		
70273383017	17	EPA 200.8	323807		
70273383018	18	EPA 200.8	323807		
70273383019	19	EPA 200.8	323807		
70273383020	20	EPA 200.8	323807		
70273383021	21	EPA 200.8	323807		
70273383022	22	EPA 200.8	323807		
70273383023	23	EPA 200.8	323807		
70273383024	24	EPA 200.8	323807		
70273383025	25	EPA 200.8	323807		
70273383026	26	EPA 200.8	323807		
70273383027	27	EPA 200.8	323807		
70273383028	28	EPA 200.8	323807		
70273383029	30	EPA 200.8	323807		
70273383030	31	EPA 200.8	323808		
70273383031	32	EPA 200.8	323808		
70273383032	33	EPA 200.8	323808		
70273383033	34	EPA 200.8	323808		
70273383034	35	EPA 200.8	323808		
70273383035	36	EPA 200.8	323808		
70273383036	37	EPA 200.8	323808		
70273383037	38	EPA 200.8	323808		
70273383038	39	EPA 200.8	323808		
70273383039	40	EPA 200.8	323808		
70273383040	41	EPA 200.8	323808		
70273383041	42	EPA 200.8	323808		
70273383042	43	EPA 200.8	323808		
70273383043	44	EPA 200.8	323808		
70273383044	45	EPA 200.8	323808		
70273383045	46	EPA 200.8	323808		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70273383046	47	EPA 200.8	323808		
70273383047	48	EPA 200.8	323808		
70273383048	49	EPA 200.8	323808		
70273383049	50	EPA 200.8	323808		
70273383050	51	EPA 200.8	323809		
70273383051	52	EPA 200.8	323809		
70273383052	53	EPA 200.8	323809		
70273383053	54	EPA 200.8	323809		
70273383054	55	EPA 200.8	323809		
70273383055	56	EPA 200.8	323809		
70273383056	57	EPA 200.8	323809		
70273383057	58	EPA 200.8	323809		
70273383058	59	EPA 200.8	323809		
70273383059	60	EPA 200.8	323809		
70273383060	61	EPA 200.8	323809		
70273383061	62	EPA 200.8	323809		
70273383062	63	EPA 200.8	323809		
70273383063	64	EPA 200.8	323809		
70273383064	65	EPA 200.8	323809		
70273383065	66	EPA 200.8	323809		
70273383066	68	EPA 200.8	323809		
70273383067	70	EPA 200.8	323809		
70273383068	71	EPA 200.8	323809		
70273383069	72	EPA 200.8	323809		
70273383070	73	EPA 200.8	323810		
70273383071	74	EPA 200.8	323810		
70273383072	75	EPA 200.8	323810		
70273383073	76	EPA 200.8	323810		
70273383074	77	EPA 200.8	323810		
70273383075	78	EPA 200.8	323810		
70273383076	79	EPA 200.8	323810		
70273383077	80	EPA 200.8	323810		
70273383078	81	EPA 200.8	323810		
70273383079	82	EPA 200.8	323810		
70273383080	83	EPA 200.8	323810		
70273383081	84	EPA 200.8	323810		
70273383082	85	EPA 200.8	323810		
70273383083	86	EPA 200.8	323810		
70273383084	87	EPA 200.8	323810		
70273383085	88	EPA 200.8	323810		
70273383086	89	EPA 200.8	323810		
70273383087	90	EPA 200.8	323810		
70273383088	91	EPA 200.8	323810		
70273383089	92	EPA 200.8	323810		
70273383090	93	EPA 200.8	323811		
70273383091	94	EPA 200.8	323811		
70273383092	95	EPA 200.8	323811		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70273383093	96	EPA 200.8	323811		
70273383094	97	EPA 200.8	323811		
70273383095	98	EPA 200.8	323811		
70273383096	99	EPA 200.8	323811		
70273383097	100	EPA 200.8	323811		
70273383099	102	EPA 200.8	323811		
70273383100	103	EPA 200.8	323811		
70273383101	104	EPA 200.8	323811		
70273383102	105	EPA 200.8	323811		
70273383103	107	EPA 200.8	323811		
70273383104	108	EPA 200.8	323811		
70273383105	109	EPA 200.8	323811		
70273383106	110	EPA 200.8	323811		
70273383107	111	EPA 200.8	323811		
70273383108	112	EPA 200.8	323811		
70273383109	113	EPA 200.8	323811		
70273383110	114	EPA 200.8	323811		
70273383111	115	EPA 200.8	323812		
70273383112	116	EPA 200.8	323812		
70273383113	117	EPA 200.8	323812		
70273383114	119	EPA 200.8	323812		
70273383115	120	EPA 200.8	323812		
70273383116	121	EPA 200.8	323812		
70273383117	122	EPA 200.8	323812		
70273383118	123	EPA 200.8	323812		
70273383119	124	EPA 200.8	323812		
70273383120	125	EPA 200.8	323812		
70273383121	126	EPA 200.8	323812		
70273383122	127	EPA 200.8	323812		
70273383123	128	EPA 200.8	323812		
70273383124	129	EPA 200.8	323812		
70273383125	130	EPA 200.8	323812		
70273383126	131	EPA 200.8	323812		
70273383128	133	EPA 200.8	323812		
70273383129	134	EPA 200.8	323812		
70273383130	135	EPA 200.8	323812		
70273383131	136	EPA 200.8	323812		
70273383132	137	EPA 200.8	323813		
70273383133	138	EPA 200.8	323813		
70273383134	139	EPA 200.8	323813		
70273383135	140	EPA 200.8	323813		
70273383136	141	EPA 200.8	323813		
70273383137	142	EPA 200.8	323813		
70273383138	143	EPA 200.8	323813		
70273383139	144	EPA 200.8	323813		
70273383140	145	EPA 200.8	323813		
70273383141	146	EPA 200.8	323813		
70273383142	147	EPA 200.8	323813		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: LEAD TESTING 10/4

Pace Project No.: 70273383

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70273383143	148	EPA 200.8	323813		
70273383144	69	EPA 200.8	323813		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Company Name: Dundee Central School District
 Street Address: 55 Water Street, Dundee, NY 14837

Contact/Report To: Schuck, Andrew
 Phone #: (607) 243-5533
 E-Mail: aschuck@dundeecs.org
 Cc E-Mail:

Customer Project #: Lead Testing

Invoice To: Sharon Scrans
 Invoice E-Mail: Scrans@dundeecs.org
 Purchase Order # (if applicable): 240236
 Quote #:

Site Collection Info/Facility ID (as applicable):

Entire School

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET New York
 Regulatory Program (DW, RCRA, etc) as applicable:

Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other
 DW PWSID # or WW Permit # as applicable:
 Field Filtered (if applicable): [] Yes [] No
 Analysis:

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Gault

Customer Sample ID	Matrk *	Comp / Grab	Collected		Res. CLZ	Composite End		Number & Type of Containers		Sample Comment
			Date	Time		Date	Time	Plastic	Glass	
1	DW G		10/4/23	3:25 AM				1		
2				3:25 AM						
3				3:29						
4				3:29						
5				3:32						
6				3:32						
7				3:33						
8				3:33						
9				3:35						
10				3:36						

Additional Instructions from Pace*:
 Collected By: Andrew Schuck
 Printed Name: Andrew Schuck
 Signature: [Signature]
 Received by/Company: [Signature] [Company]
 Date/Time: 10/4/23 9:00
 Tracking Number:
 Delivered by: [] In-Person [] Courier
 [] FedEx [] UPS [] Other
 Page: 1 of 1

LAB USE ONLY - Affix Worker's Login Label Here
WO#: 70273383

70273383

Specify Container Size **
 ** Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL vial, (7) EnCore, (8) TerraCore, (9) Other
 Identify Container Preservative Type***
 *** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Analysis Requested
 Proj. Mgr: Daniel Bonitto
 AcctNum / Client ID:
 Table #: 8628
 Profile / Template:
 Prelog / Bottle Ord. ID: 1145598
 Sample Comment

Preservation non-conformance identified for sample

CHAIN-OF-CUSTODY Analytical Request Document
 Chain of Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY - Affix Workorder/Login Label Here



Scan QR Code for instructions

Company Name: **Dundee Central School District**
 Street Address: **55 Water Street, Dundee, NY 14837**
 Customer Project #: **Lead Testing**
 Project Name: **Entire School**
 Site Collection Info/Facility ID (if applicable):
 Invoice To: **Sharon Crans**
 Invoice E-Mail: **scrans@dundeecs.org**
 Purchase Order # (if applicable): **240236**
 Quote #:

Contact/Report To: **Schuck, Andrew**
 Phone #: **(607) 243-5533**
 E-Mail: **aschuck@dundeecs.org**
 Cc E-Mail:

Specify Container Size **
 Identify Container Preservative Type***
 Analysis Requested

** Container Size: (1) 1L (2) 500mL (3) 350mL (4) 250mL (5) 100mL (6) 40mL vol. (7) Encore, (8) TerraCore, (9) Other
 *** Preservative Type: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sulf. Thiosulfate, (9) Ascorbic Acid, (10) NaOH, (11) Other

Time Zone Collected: () AK () PT () MT () CT () ET
 Regulatory Program (DW, RCRA, etc) as applicable: **New York**
 Rush (pre-approval required):
 () 2 Day () 3 day () 5 day () Other
 Date Results Requested:
 Field Filtered (if applicable): () Yes () No
 Analysis:

Proj. Mgr:
Daniel Bonitto
 Account / Client ID:
 Table #:
 Profile / Template:
862B
 Prelog / Bottle Ord. ID:
1145598
 Sample Comment:
 Preservation non-compliance identified for sample.

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Res. CL2	Number & Type of Container
			Date	Time		
11	DW	G	10/4/23	3:36 AM		1
12				3:37		
13				3:37		
14				3:39		
15				3:39		
16				3:42		
17				3:43		
18				3:45		
19				3:46		
20				3:51		

Additional Instructions from Pace:	Thermometer ID:	Obs. Temp. [C]	Corrected Temp. [C]

Collected By: **Andrew Schuck**
 Printed Name:
 Signature: *Andrew Schuck*
 Received by/Company (Signature): *[Signature]*
 Date/Time: **10/9/23 9:20**
 Tracking Number:
 Delivered by: () In-Person () Courier () FedEx () UPS () Other
 Page: **1** of **1**

Customer Remarks / Special Conditions / Possible Hazards:
 Submitting a sample via this chain of custody constitutes acknowledgment and acceptance of the Pace® Terms and Conditions found at <https://www.pacelabs.com/resource-library/resource/pace-terms-and-conditions/>



Scan QR Code for instructions

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Pace* Location requested (City/State):
Pace Analytical Long Island NY
575 Great Hollow Rd, Molville, NY 11747

Company Name: Dundee Central School District
Street Address: 55 Water Street, Dundee, NY 14837

Customer Project #: Lead Testing
Project Name: Entire School

Site Collection Info/Facility ID (as applicable):
Entire School

Contact/Report To: Schuck, Andrew
Phone #: (607) 243-5533
E-Mail: aschuck@dundeecs.org
Cc E-Mail:

Invoice To: Sharon Crans
Invoice E-Mail: Scrans@dundeecs.org
Purchase Order # (if applicable): 240236
Quote #:

County / State origin of sample(s): New York

Regulatory Program (DW, RCRA, etc.) as applicable:
Rush (pre-approval required):
[] 2 Day [] 3 day [] 5 day [] Other
Date Results Requested:

Field Filtered (if applicable): [] Yes [] No
Analysis:

* Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk
Other (OT),

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Composite End	Res. CL2	Number & Type of Container	
			Date	Time			Date	Time
21	DW	G	10/4/23	3:52 AM				1
22				3:52				
23				3:55				
24				3:56				
25				3:56				
26				3:58				
27				3:59				
28				3:59				
29				—				
30				4:03				

Customer Remarks / Special Conditions / Possible Hazards:

Collected By: Andrew Schuck
Printed Name: Andrew Schuck
Signature: [Signature]

Relinquished by/Company (Signature)
Relinquished by/Company (Signature)
Relinquished by/Company (Signature)
Relinquished by/Company (Signature)

Date/Time:
Date/Time:
Date/Time:
Date/Time:

Received by/Company (Signature)
Received by/Company (Signature)
Received by/Company (Signature)
Received by/Company (Signature)

Date/Time:
Date/Time:
Date/Time:
Date/Time:

Tracking Number:
Delivered by: [] In Person [] Courier
[] FedEx [] UPS [] Other

Company Name: Dundee Central School District
 Street Address: 55 Water Street, Dundee, NY 14837
 Customer Project #: Lead Testing
 Project Name: Entire School
 Site Collection Info/Facility ID (as applicable):
 Entire School

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields


Contact/Report To: Schuck, Andrew
 Phone #: (607) 243-5533
 E-Mail: aschuck@dundeecs.org
 Cc E-Mail:
 Invoice To: Sharon Crans
 Invoice E-Mail: socrans@dundeecs.org
 Purchase Order # (if applicable): 240236
 Quote #:

County / State origin of sample(s): New York
 Regulatory program (DW, RCRA, etc.) as applicable:
 Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other
 Date Results Requested:
 DW PMSID # or WV Permit # as applicable:
 Field Filtered (if applicable): [] Yes [] No
 Analysis:

Customer Sample ID	Comp / Grab	Matrix *	Collected (or Composite Start)		Composite End		Res. CL2	Number & Type of Container
			Date	Time	Date	Time		
31	DW G		10/4/23	4:04 AM			1	plastic
32				4:04				
33				4:05				
34				4:07				
35				4:07				
36				4:10				
37				4:10				
38				4:12				
39				4:12				
40				4:14				

Customer Remarks / Special Conditions / Possible Hazards:
 Collected By: Andrew Schuck
 Printed Name: Andrew Schuck
 Signature: *Andrew Schuck*
 Received by/Company: *[Signature]*
 Date/Time: 10/4/23 9:00
 Received by/Company: *[Signature]*
 Date/Time:
 Received by/Company: *[Signature]*
 Date/Time:
 Received by/Company: *[Signature]*
 Date/Time:

LAB USE ONLY - Affix Workorder/Label Here



Scan QR Code for instructions

Specify Container Size **
 *** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sulf. Thioacetate, (9) Acetic Acid, (10) HEDTA, (11) Other
 Analysis Requested

Proj. Mgr:
 Daniel Bonitto
 Accrual / Client ID:
 Table #:
 Profile / Template:
 862B
 Pricing / Bottle Ord. ID:
 1145598

Lab Use Only

Preservation non-performance identified for

Sample Comment

Additional Instructions from Pace:
 Thermometer ID: Obs. Temp. [C] Corrected Temp. [C]
 Correction Factor [C]:
 Tracking Number:
 Delivered by: [] In-Person [] Courier
 [] FedEx [] UPS [] Other
 Page: 1 of 1

Company Name: Dundee Central School District
 Street Address: 55 Water Street, Dundee, NY 11937

Customer Project #: Lead Testing
 Project Name: Entire School

Site Collection info/Facility ID (if applicable):
 Entire School

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET
 Date Deliverables: [] Level II [] Level III [] Level IV
 [] EQUIS [] Other

Regulatory Program (DW, RCRA, etc.) as applicable:
 Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other
 Date Results Requested:

* Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (O), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (O1), Surface Water (SW), Sediment (SED), Sludge (S), Cask

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Contact/Report To: Schuck, Andrew
 Phone #: (607) 243-5533
 E-Mail: aschuck@dundeecs.org
 Cc E-Mail:

Invoice To: Sharon Crans
 Invoice E-Mail: Scrans@dundeecs.org
 Purchase Order # (if applicable): 240236
 Quote #:

County / State origin of sample(s): New York

DW PWSID if or WW Permit # as applicable:
 Field Filtered (if applicable): [] Yes [] No
 Analysis:

Matrix	Customer Sample ID	Comp / Grab	Collected (or Composite Start)		Composite End Date	Time	Res. CL2	Number & Type of Containers	
			Date	Time				Plastic	Glass
DW G			10/4/23	4:14 AM				1	
				4:15					
				4:16					
				4:17					
				4:17					
				4:18					
				4:22					
				4:22					
				4:22					

Customer Remarks / Special Conditions / Possible Hazards:

Collected By: Andrew Schuck
 Printed Name: Andrew Schuck
 Signature: *Andrew Schuck*

Received by/Company (Signature) *ARACI*
 Date/Time: 10/4/23 9:10
 Received by/Company (Signature)
 Date/Time:
 Received by/Company (Signature)
 Date/Time:
 Received by/Company (Signature)
 Date/Time:

Additional Instructions from Pace:
 Thermometer ID: Correction Factor (°C): Obs. Temp. (°C) Corrected Temp. (°C)
 Tracking Number:
 Delivered by: [] In-Person [] Courier
 [] FedEx [] UPS [] Other
 Page: 1 of 1



Scan QR Code for instructions

Specify Container Size **
 Identify Container Preservative Type**
 Analysis Requested

Lab Use Only
 Proj. Mgr: Daniel Bonitto
 Account / Client ID:
 Table #: 8628
 Profile / Template:
 Prelog / Bottle Ord. ID: 1145598
 Sample Comment

**Container Size: (1) 1L (2) 500mL (3) 250mL (4) 125mL (5) 100mL (6) 40mL vol. (7) Ericore (8) TerraCure (9) Other
 *** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl (5) H2O2, (6) Zn acetate, (7) Nitric, (8) Sul. Thiourea, (9) Acetic Acid, (10) Media, (11) Other

Preservation non-performance identified for sample.

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Company Name: Dundee Central School District
 Street Address: 55 Water Street, Dundee, NY 14837

Contact/Report To: Schuck, Andrew
 Phone #: (607) 243-5533
 E-Mail: aschuck@dundeecs.org
 Cc E-Mail:

Customer Project #: Lead Testing
 Project Name: Entire School

Site Collection Info/Facility ID (as applicable):
 Entire School

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET
 Regulatory Program (DW, RCRA, etc.) as applicable: New York

Date Deliverables:
 [] Level II [] Level III [] Level IV
 [] EQUIS
 [] Other

Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other
 DW PWSID # or WW Permit # as applicable:
 DW PWSID # or WW Permit # as applicable:
 Field Filtered (if applicable): [] Yes [] No
 Analysis Requested:

* Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Sediment (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Cawk

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Composite End		Res. CL2	Number & Type of Container		200g (Lead)
			Date	Time	Date	Time		plastic	Glass	
51	DW G		10/4/23	4:22 AM					7	
52				4:24						
53				4:24						
54				4:26						
55				4:26						
56				4:27						
57				4:27						
58				4:28						
59				4:28						
60				4:29						

Customer Remarks / Special Conditions / Possible Hazards:

Collected By: Andrew Schuck
 Printed Name: Andrew Schuck
 Signature: *Andrew Schuck*

Received by/Company (Signature): *Ch. J. [Signature]*
 Date/Time: 10/9/23 9:00

Received by/Company (Signature): *SPACEY*
 Date/Time:

Received by/Company (Signature):
 Date/Time:

Received by/Company (Signature):
 Date/Time:

Received by/Company (Signature):
 Date/Time:

Tracking Number:
 Delivered by: [] In Person [] Courier
 [] FedEx [] UPS [] Other

Page: 1 of 1



Scan QR Code for instructions

Specify Container Size **
 15mL (1) 10mL (6) 40mL vol. (7) EnCore (8) TerraCore (9) Other

Identify Container Preservative Type***
 *** Preservative Types (1) None, (2) HClO2, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sulf. Thiosulfate, (9) Ascorbic Acid, (10) NaOH, (11) Other

Analysis Requested

Lab Use Only
 Proj. Mgr:
 Daniel Bonitto
 Account / Client ID:
 Table #:
 Profile / Template:
 8628
 Prelog / Bottle Ord. ID:
 1145598

Preservation non-compliance identified for

Additional Instructions from Pace*:
 # Coolers: Thermometer ID: Obs. Temp (°C) Corrected Temp (°C)

Company Name: Dundee Central School District
 Street Address: 55 Water Street, Dundee, NY 11837

Customer Project #: Lead Testing

Project Name: Entire School

Site Collection Info/Facility ID (as applicable):
 Entire School

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET
 Data Deliverables:
 [] Level II [] Level III [] Level IV
 [] EQUIS
 [] Other:

Regulatory Program (DW, RCRA, etc.) as applicable:
 Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other:
 Date Results Requested:

Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Sludge (S), Soil
 Other (OI), Surface Water (SW), Sediment (SED), Sludge (S), Soil

Customer/Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start) Date	Time	Composite End Date	Time	Res. CL2	Number & Type of Container
61	DW	G	10/4/23	4:29 AM				1
62				4:30				
63				4:30				
64				4:31				
65				4:31				
66				4:33				
67	Not Needed							
68				4:34				
69				4:34				
70				4:36				

Customer Remarks / Special Conditions / Possible Hazards:

Collected By: Andrew Schuck
 Printed Name: Andrew Schuck
 Signature: *Andrew Schuck*

Date/Time: *10/4/23 9:20*
 Date/Time: *10/4/23 9:20*
 Date/Time:
 Date/Time:

Received by/Company (Signature): *Andrew Schuck*
 Received by/Company (Signature):
 Received by/Company (Signature):
 Received by/Company (Signature):

Tracking Number:
 Delivered by: [] In-Person [] Courier
 [] FedEx [] UPS [] Other
 Page: 1 of 1

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Contact/Report To: Schuck, Andrew
 Phone #: (607) 243-5533
 E-Mail: aschuck@dundeecs.org
 Cc E-Mail:

Invoice To: Sharon Crans
 Invoice E-Mail: Scrans@dundeecs.org
 Purchase Order # (if applicable): 240236
 Quote #:

County / State origin of sample(s): New York

Specify Container Size:
 Identify Container Preservative Type:
 Analyte(s) Requested:

Field Filtered (if applicable): [] Yes [] No
 Analysis:
 DW (PWSID) or WW Permit # as applicable:
 Field Filtered (if applicable): [] Yes [] No

Preservation non-performance identified for	Proj. Mgr:	Table #:	Profile / Template:	Prelog / Bottle Ord. ID:	Sample Comment
	Daniel Bonitto		8628	1145598	

Additional Instructions from Pace:
 # Colvets: Thermometer ID: Correction Factor (°C): Obs. Temp (°C) Corrected Temp. (°C)
 Date/Time: 10/4/23 9:20
 Date/Time: 10/4/23 9:20
 Date/Time:
 Date/Time:

LAB USE ONLY - After Workorder Login Label Here

WO#: 70273383



**Container Size: (1) 1L (2) 500mL (3) 250mL (4) 125mL (5) 100mL (6) 40mL vol. (7) EnCore, (8) TerraCore, (9) Other
 *** Preservative Types: (1) None, (2) HClO3, (3) H2SO4, (4) HCl, (5) HNO3, (6) No Preservative, (7) Nitric Acid, (8) Sulphuric Acid, (9) Acetic Acid, (10) Nitric Acid, (11) Other

Lab Use Only	Proj. Mgr:	Table #:	Profile / Template:	Prelog / Bottle Ord. ID:	Sample Comment
	Daniel Bonitto		8628	1145598	

Company Name: Dundee Central School District
 Street Address: 55 Water Street, Dundee, NY 14837
 Customer Project #: Lead Testing
 Project Name: Entire School
 Site Collection Info/Facility ID (as applicable):

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields.

Contact/Report To: Schuck, Andrew
 Phone #: (607) 243-5533
 E-Mail: aschuck@dundeecs.org
 Cc E-Mail:
 Invoice To: Sharon Crans
 Invoice E-Mail: Scrans@dundeecs.org
 Purchase Order # (if applicable): 240236
 Quote #:

County / State origin of sample(s): New York
 Regulatory Program (DW, RCRA, etc.) is applicable:
 Rush (pre-approval required):
 2 Day 3 day 5 day Other
 Date Results Requested:
 DW PWSID # or WW Permitt # is applicable:
 Field Filtered (if applicable): Yes No
 Analysis:

* Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Surface Water (SW), Sediment (SED), Sludge (SL), Crank
 Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Crank

Matrix	Comp/Grab	Collected (or Composite Start)		Composite End		Res. CL2	Number & Type of Containers
		Date	Time	Date	Time		
DW	G	10/4/23	4:36 AM				1
			4:37				
			4:38				
			4:38				
			4:39				
			4:39				
			4:40				
			4:41				
			4:41				
			4:42				

Customer Remarks / Special Conditions / Possible Hazards:
 Collected By: Andrew Schuck
 Printed Name: Andrew Schuck
 Signature: *Andrew Schuck*
 Received by/Company: *ASCHUCK*
 Signature: *ASCHUCK*
 Received by/Company: *ASCHUCK*
 Signature: *ASCHUCK*
 Received by/Company: *ASCHUCK*
 Signature: *ASCHUCK*
 Received by/Company: *ASCHUCK*
 Signature: *ASCHUCK*

LAB USE ONLY - Affix Workorder/Log-in Label Here



Scan QR Code for instructions

Specify Container Size **
 Identify Container Preservative Type***
 Analysis Requested

Proj. Mfr:
 Daniel Bonitto
 Account / Client ID:
 Table ID:
 Profile / Template:
 8628
 Prelog / Bottle Ord. ID:
 1145598
 Sample Comment

Container Size	Preservative Type	Analysis Requested	Proj. Mfr	Account / Client ID	Table ID	Profile / Template	Prelog / Bottle Ord. ID	Sample Comment

Additional Instructions (from Page*):

Coolers: Thermometer ID: Correction Factor (°C): Obs. Temp (°C) Corrected Temp (°C)
 Tracking Number:
 Date/Time: 10/9/23 9:00
 Date/Time:
 Date/Time:
 Date/Time:
 Delivered by: In-Person Courier
 FedEx UPS Other
 Page: 1 of 1

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Pace* Location Requested (City/State):
 Pace Analytical Long Island NY
 575 Broad Hollow Rd, Melville, NY 11747

Company Name: Dundee Central School District
 Street Address: 55 Water Street, Dundee, NY 14837

Customer Project #: Lead Testing

Project Name: Entire School

Site Collection Info/Facility ID (as applicable):

Contact/Report To: Schwack, Andrew
 Phone #: (607) 243-5533
 E-Mail: aschwack@dundees.org
 Cc E-Mail:

Invoice To: Sharon Crans
 Invoice E-Mail: scrans@dundees.org
 Purchase Order # (if applicable): 240236
 Quote #:

County / State origin of sample(s): New York

Regulatory Program (DW, RCRA, etc.) as applicable:

Rush (Pre-approval required):
 2 Day 3 day 5 day Other

Date Results Requested:

* Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Urinassay (U), Vapor (V), Other (OT), Surface Water (SW), Sediment (SD), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Collected (or Composite Start)		Composite End		Res. CL2	Number & Type of Containers	
		Date	Time	Date	Time		PLASTIC	GLASS
81	DW G	10/4/23	4:42 AM				1	
82			4:44					
83			4:44					
84			4:47					
85			4:47					
86			4:47					
87			4:50					
88			4:52					
89			4:52					
90			4:55					

Customer Remarks / Special Conditions / Possible Hazards:

Collected By: Andrew Schwack
 Printed Name: Andrew Schwack
 Signature: *Andrew Schwack*

Requested by/Company: (Signature)

Received by/Company: (Signature)

Requested by/Company: (Signature)

Received by/Company: (Signature)

Date/Time:

Date/Time:

Date/Time:

Date/Time:

Requested by/Company: (Signature)

Received by/Company: (Signature)

Requested by/Company: (Signature)

Received by/Company: (Signature)

Date/Time:

Date/Time:

Date/Time:

Date/Time:

Tracking Number: 10941894:0

Delivered by: In Person Courier

FedEx UPS Other

Page: 1 of 1

LAB USE ONLY - Affix Workorder/Log/Lab Here



Scan QR Code for instructions

Specify Container Size **

Identify Container Preservative Type***

Analysis Requested

Proj. Mgr:	Daniel Bonitto
Acetium / Client ID:	
Tablet #:	
Profile / Template:	8628
Prelog / Bottle Ord. ID:	1145598
Sample Comment:	

Preservation non-conformance identified for sample.

Additional Instructions from Page*:

Casks: Thermometer ID: Correction Factor (°C): Obs. Temp. (°C) Corrected Temp. (°C)

Company Name: Dundee Central School District
Street Address: 55 Water Street, Dundee, NY 14837

Customer Project #: Lead Testing
Project Name: Entire School

Site Collection info/Facility ID (as applicable):
Entire School

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET

Date Deliverables: [] Level II [] Level III [] Level IV

[] EQUIS [] Other

Regulatory Program (DW, RCRA, etc.) as applicable: New York

Rush (Pre-approval required): [] 2 Day [] 3 day [] 5 day [] Other

Invoice E-Mail: aschuck@dundeecs.org

Field Filtered (if applicable): [] Yes [] No

Analysis: [] Yes [] No

Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Surface Water (SW), Sediment (SED), Sludge (SL), Gunk

Other (OI), Surface Water (SW), Sediment (SED), Sludge (SL), Gunk

Customer Sample ID

Matrix *

Comp / Grab

Collected (or Composite Start) Date Time

Composite End Date Time

Res. CL2

Number & Type of Containers

Plastic

Glass

91 DW G 10/4/23 4:55 AM

92 4:58

93 4:58

94 4:58

95 5:50

96 5:50

97 5:51

98 5:52

99 5:53

100 6:05

Customer Remarks / Special Conditions / Possible Hazards:

Collected By: Andrew Schuck

Printed Name: Andrew Schuck

Signature: [Signature]

Received by/Company (Signature)

Received by/Company (Signature)

Received by/Company (Signature)

Received by/Company (Signature)

Date/Time

Date/Time

Date/Time

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Contact/Report To: Schuck, Andrew

Phone #: (607) 243-5533

E-Mail: aschuck@dundeecs.org

Cc E-Mail:

Invoice To: Sharon Crans

Invoice E-Mail: Scrans@dundeecs.org

240236

Purchase Order # (if applicable):

Quote #:

County / State origin of sample(s): New York

Regulatory Program (DW, RCRA, etc.) as applicable:

Rush (Pre-approval required):

[] 2 Day [] 3 day [] 5 day [] Other

Invoice E-Mail:

Field Filtered (if applicable): [] Yes [] No

Analysis:

[] Yes [] No

Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Surface Water (SW), Sediment (SED), Sludge (SL), Gunk

Other (OI), Surface Water (SW), Sediment (SED), Sludge (SL), Gunk

Customer Sample ID

Matrix *

Comp / Grab

Collected (or Composite Start) Date Time

Composite End Date Time

Res. CL2

Number & Type of Containers

Plastic

Glass

91 DW G 10/4/23 4:55 AM

92 4:58

93 4:58

94 4:58

95 5:50

96 5:50

97 5:51

98 5:52

99 5:53

100 6:05

Customer Remarks / Special Conditions / Possible Hazards:

Collected By: Andrew Schuck

Printed Name: Andrew Schuck

Signature: [Signature]

Received by/Company (Signature)

Received by/Company (Signature)

Received by/Company (Signature)

Received by/Company (Signature)

Date/Time

Date/Time

Date/Time

LAB USE ONLY - Affix Workorder/LogIn Label Here



Scan QR Code for instructions

Specify Container Size **

Identify Container Preservative Type**

Analysis Requested

Proj. Mgr:

Daniel Bonitto

Archium / Client ID:

Table #:

8628

Profile / Template:

1145598

Project / Bottle Ord. ID:

1145598

Sample Comment

Preservation non-performance identified for sample.

Lab Use Only

Tracking Number:

Delivered by: [] In-Person [] Courier

[] FedEx [] UPS [] Other

Page: 1 of 1

ENV-FRM-CORQ-0019_v01_082123 ©

Company Name: Dundee Central School District
 Street Address: 55 Water Street, Dundee, NY 14837

Customer Project #: Lead Testing
 Project Name: Entire School

Site Collection Info/Facility ID (as applicable):
 Entire School

Time Zone Collected: [] JAK [] PT [] MT [] CT [] ET

Data Deliverables:
 [] Level II [] Level III [] Level IV
 [] EQUUS
 [] Other

Regulatory Program (DW, RCRA, etc.) as applicable:
 Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other

Date Results Requested:
 DW PWSID # or WW Permit # as applicable:
 Analysis:
 Field Filtered (if applicable): [] Yes [] No

Country / State origin of sample(s): New York

Contact/Report To: Schuck, Andrew
 Phone #: (607) 243-5533
 E-Mail: aschuck@dundeecs.org
 Cc E-Mail:

Invoice To: Sharon Crans
 Invoice E-Mail: Scrans@dundeecs.org
 Purchase Order # (if applicable): 240236
 Quote #:

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (O), Surface Water (SW), Sediment (SED), Sludge (SL), Cask

Matrix *	Customer Sample ID	Collected (or Composite Start) Date	Time	Composite End Date	Time	Res. CL2	Number & Type of Containers	
							Plastic	Glass
DW G	104/23	5:12 AM					1	
		5:12						
		5:12						
		5:15						
		5:15						
		5:15						
		5:15						
		5:19						
		5:19						

Additional Instructions (from Pace):
 # Coolers: Thermometer ID: Correction Factor (°C): Obs. Temp (°C) Corrected Temp (°C)
 Collected By: Andrew Schuck
 Printed Name: Andrew Schuck
 Signature: [Signature]
 Received by/Company: [Signature] RACTU
 Received by/Company: [Signature]
 Received by/Company: [Signature]
 Received by/Company: [Signature]

LAB USE ONLY - AMX Workorder/Login Label Here

Specify Container Size **

Identify Container Preservative Type***

Analysis Requested

Profil. Mgr:
 Daniel Benitto
 Accr Num / Client ID:
 Table #:
 Profile / Template:
 8628
 Prelog / Bottle Ord. ID:
 1145598

Sample Comment

Preservation non-compliance identified for sample.

** Container Size: (1) 1L (2) 500mL (3) 250mL (4) 125mL (5) 100mL (6) 40mL vol. (7) EnCore, (8) TerraCore, (9) Other
 *** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) H2O2, (6) Zn Acetate, (7) NaHSO4, (8) Sulf. Thiosulfate, (9) Ascorbic Acid, (10) Starch, (11) Other

Scan QR Code for instructions

Tracking Number:
 Date/Time: 10/9/23 9:20
 Date/Time:
 Date/Time:
 Date/Time:

Delivered by: [] In Person [] Courier
 [] FedEx [] UPS [] Other

Page: 1 of 1

ENV-FRM-CORQ-0019_v01_082123 ©

Chain-of-Custody Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Company Name: Dundee Central School District
 Street Address: 55 Water Street, Dundee, NY 14837

Contact/Report To: Schuck, Andrew
 Phone #: (607) 243-5533
 E-Mail: aschuck@dundeecs.org
 Cc E-Mail:

Customer Project #: _____
 Project Name: Lead Testing

Site Collection Info/Facility ID (As Applicable):
 Entire School

Invoice To: Sharon Crans
 Invoice E-Mail: Scrans@dundeecs.org
 Purchase Order # (if applicable): 240236
 Quote #:

Time Zone Collected: [] JAX [] MT [] CT [] ET New York

Data Deliverables:
 [] Level II [] Level III [] Level IV
 [] EDUS
 [] Other: _____

Regulatory Program (DW, RCRA, etc.) as applicable: _____
 DW PWSID # or WW Permit # as applicable: _____
 Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other: _____
 Date Results Requested: _____
 Field Filtered (if applicable): [] Yes [] No
 Analysis: _____

* Matrix Codes: (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (O), Wipe (WI), Tissue (TS), Urinassay (U), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Chalk

Customer Sample ID	Matrix *	Comp / Grab	Collected (for Composite Start) Date Time	Composite End Date Time	Res. CL2	Number & Type of Container
121	DW	G	10/4/23 5:20AM			1
122			5:21			
123			5:21			
124			5:23			
125			5:23			
126			5:25			
127			5:25			
128			5:27			
129			5:27			
130			5:27			

Additional Instructions from Pace[®]:

Collected By: Andrew Schuck
 Printed Name: Andrew Schuck
 Signature: *Andrew Schuck*

Received by/Company: (Signature) *[Signature]*
 Date/Time: 10/4/23 9:00

Received by/Company: (Signature) _____
 Date/Time: _____

Received by/Company: (Signature) _____
 Date/Time: _____

Received by/Company: (Signature) _____
 Date/Time: _____

Tracking Number: _____
 Delivered by: [] In-Person [] Courier
 [] FedEx [] UPS [] Other

Page: 1 of 1



Scan QR Code for instructions

Specify Container Size **

Identify Container Preservative Type***

Analysis requested

Lab Use Only

Proj. Mgr: Daniel Bonitto
 Acct/Num / Client ID:
 Table #:
 Profile / Template: 8628
 Protog. / Bottle Ord. ID: 1145598

Sample Comment

**Container Size: (1) 1L (2) 500mL (3) 250mL (4) 125mL (5) 100mL (6) 40mL vial, (7) EnCore, (8) TerraCure, (9) Other

*** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) H2O2, (6) Zn Acetate, (7) NaHSO4, (8) Sud. Thiosulfate, (9) Ascorbic Acid, (10) StcHCl, (11) Other

Preservative performance identified for sample.



LAB USE ONLY - Mix Workorder Login Label Here

Scan QR Code for instructions

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Pace® Location requested (City/State): Pace Analytical Long Island NY
575 Broad Hollow Rd, Melville, NY 11747

Company Name: Dundee Central School District
Street Address: 55 Water Street, Dundee, NY 14837

Contact/Report To: Schuck, Andrew
Phone #: (607) 243-5533
E-Mail: aschuck@dundeecs.org
Cc E-Mail:

Invoice To: Sharon Crans
Invoice E-Mail: Scrans@dundeecs.org
Purchase Order # (if applicable): 240236
Quote #:

Customer Project #:
Project Name: Lead Testing
Site Collection Info/Facility ID (as applicable): Entire School

Time Zone Collected: [] JAK [] JPT [] JMT [] JCT [] ET
Regulatory Program (DWR, RCRA, etc.) as applicable: New York

Date Deliverable:
[] Level II [] Level III [] Level IV
[] EQUUS
[] Other

Rush (Pre-approval required): DW PMSID # or WV Permit # as applicable:
[] 2 Day [] 3 day [] 5 day [] Other
Date Results Requested:

Field Filtered (if applicable): [] Yes [] No
Analysis:

* Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Biossary (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Gauk

Customer Sample ID	Mark	Comp / Grab	Collected (or Composite Start)		Composite End		Res. CL2	Number & Type of Containers	
			Date	Time	Date	Time		plastic	Glass
131	DW G		10/4/23	5:28 AM				1	
132				5:28					
133				5:30					
134				5:30					
135				5:31					
136				5:31					
137				5:32					
138				5:33					
139				5:33					
140				5:34					

Additional instructions from Pace®:

Collected By: Andrew Schuck
Principed Name: Andrew Schuck
Signature: *Andrew Schuck*


Received by/Company (Signature): *[Signature]*
Date/Time: 10/4/23 9:00

Tracking Numbers:

Delivered by: [] In Person [] Courier
[] FedEx [] UPS [] Other

Page: 1 of 1

Submitting a sample via this chain of custody constitutes acknowledgment and acceptance of the Pace® Terms and Conditions found at <https://www.pacelabs.com/resource-library/resource/pace-terms-and-conditions/>



Pace[®] Location Requested (City/State):
 Pace Analytical Long Island NY
 575 Broad Hollow Rd, Malville, NY 11747

Company Name: Dundee Central School District
 Street Address: 55 Water Street, Dundee, NY 14837

Customer Project #: Lead Testing

Site Collection Info/Facility ID (as applicable):
 Entire School

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET

Data Deliverables:
 [] Level II [] Level III [] Level IV
 [] EQUIS
 [] Other

Regulatory Program (DW, RCRA, etc.) as applicable:
 DW PWSID # or NY Permit # as applicable:
 Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other
 Date Results Requested:

* Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk
 Other (OT),

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Contact/Report To: Schuck, Andrew
 Phone #: (607) 243-5533
 E-Mail: aschuck@dundeecs.org
 Cc E-Mail:

Invoice To: Sharon Grans
 Invoice E-Mail: Sgrans@dundeecs.org
 Purchase Order # (if applicable): 240236
 Quote #:

Country / State origin of sample(s): New York

Field Filtered (if applicable): [] Yes [] No
 Analysis:

Matrix	Comp / Grab	Collected (or Composite Start) Date	Time	Composite End Date	Time	Res. CL2	Number & Type of Containers	
							Plastic	Glass
DW G		10/4/23	5:34 AM				1	
			5:35					
			5:36					
			5:37					
			5:38					
			5:39					
			5:40					

Customer Remarks / Special Conditions / Possible Hazards:
 Collected By: *Andrew Schuck*
 Printed Name: Andrew Schuck
 Signature: *Andrew Schuck*
 Received by/Company (Signature): *Sharon Grans*
 Date/Time: 10/4/23 7:50
 Received by/Company (Signature):
 Date/Time:
 Received by/Company (Signature):
 Date/Time:
 Received by/Company (Signature):
 Date/Time:



LAB USE ONLY: Mix Workorder/Login Label Here

Specify Container Size **
 Identify Container Preservative Type***
 Analysis Requested

**Container Size: (1) 1L (2) 500mL (3) 250mL (4) 125mL (5) 100mL (6) 40mL vol. (7) Encore, (8) Ten-A-Cure, (9) Other
 *** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) H3PO4, (6) Zn Acetate, (7) NaOH, (8) Sulf. Thiourea, (9) Acetic Acid, (10) MDCR, (11) Other

Lab Use Only
 Proj. Mgr: Daniel Bonitto
 Account / Client ID:
 Table #:
 Profile / Template: 8628
 Protocol / Bottle Ord. ID: 1145598
 Sample Comment

Preservative non-performance identified for

Additional Instructions from Pace:
 # Colvars: Thermometer ID: Correction Factor (°C): Obs. Temp. (°C) Corrected Temp. (°C)

WO#: 70273383
PM: DHB **Due Date: 10/23/23**
CLIENT: DCSD

Client Name: Dundee Central School District Project #
 Courier: Fed Ex UPS USPS Client Commercial Pace Other
 Tracking #: 7054 7899 0079

Custody Seal on Cooler/Box Present: Yes No Seals intact: Yes No Temperature Blank Present: Yes No
 Packing Material: Bubble Wrap Bubble Bags Ziploc None Other Type of Ice: Wet Blue None
 Thermometer Used: TH198 Correction Factor: +0.3 Samples on ice, cooling process has begun
 Cooler Temperature (°C): 20.7 Cooler Temperature Corrected (°C): 30.0 Date/Time 5035A kits placed in freezer _____
 Temp should be above freezing to 6 °C

USDA Regulated Soil (N/A, water sample)
 Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? Yes No

Did samples originate from a foreign source including Hawaii and Puerto Rico)? Yes No

If Yes to either question, fill out a Regulated Soil Checklist (ENV-FRM-MELV-0076) and include with SCUR/COC paperwork.

Date and Initials of person examining contents: A.S. 10/10/23

	COMMENTS:
Chain of Custody Present <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	1.
Chain of Custody Filled Out: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	2.
Chain of Custody Relinquished: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	3.
Sampler Name & Signature on COC: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	4.
Samples Arrived within Hold Time: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	5.
Short Hold Time Analysis (<72hr): <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	6.
Rush Turn Around Time Requested <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	7.
Sufficient Volume: (Triple volume provided for MS/MSD) <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	8.
Correct Containers Used: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	9.
-Pace Containers Used: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	
Containers Intact: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	10.
Filtered volume received for Dissolved tests <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	11. Note: if sediment is visible in the dissolved container.
Sample Labels match COC: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	12.
-Includes date/time/ID/Analysis: Matrix: SL WT OIL <u>OTHER</u>	

Date and Initials of person checking preservation: A.S. 10/10/23

All containers needing preservation have been pH paper Lot # <u>HC 208072</u> All containers needing preservation are found to be in compliance with method recommendation? (HNO ₃ , H ₂ SO ₄ , HCl, NaOH > 9 Sulfide, <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A NAOH > 12 Cyanide) Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water). Per Method, VOA pH is checked after analysis	13. <input type="checkbox"/> HNO ₃ <input type="checkbox"/> H ₂ SO ₄ <input type="checkbox"/> NaOH <input type="checkbox"/> HCl Sample #
	Initial when completed: Lot # of added preservative: Date/Time preservative added:
Samples checked for dechlorination. <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A KI starch test strips Lot #	14. Positive for Res., Chlorine? Y N
Residual chlorine strips Lot #	15. Positive for Sulfide? Y N
SM 4500 CN samples checked for sul <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	
Lead Acetate Strips Lot #	
Headspace in VOA Vials (>6mm) <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	16.
Trip Blank Present: <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	17.
Trip Blank Custody Seals Present <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	

Client Notification/ Resolution: _____ Field Data Required? Y / N
 Person Contacted: _____ Date/Time: _____
 Comments/ Resolution: _____

* PM (Project Manager) review is documented electronically in LIMS.